

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003821  
11:49 15/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:14  
PRESSURIZING 1:20  
STERILIZING 2:12  
136.3 C 222kPa 2:12  
136.2 C 223kPa 2:42  
136.3 C 223kPa 3:12  
136.3 C 222kPa 3:42  
136.2 C 223kPa 4:12  
136.1 C 219kPa 4:42  
136.3 C 222kPa 5:12  
136.4 C 222kPa 5:42  
136.4 C 222kPa 5:42  
VENTING 5:42  
AIR DRYING 7:52  
CYCLE COMPLETE 11:00

-----  
Digital Signature #  
6FE3644E8F5E2FE3  
-----