
Midmark M11 - v1.0.5
Total Cycles: 644

Name: RITTER2
Sterilizer ID: V2313228
12/16/2021 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.4 F 0.0
2:00 175.1 F 3.4
4:00 186.8 F 7.6
6:00 207.9 F 12.4
8:00 223.7 F 6.8
10:00 235.3 F 11.8
12:00 235.0 F 9.3
14:00 238.6 F 10.0
16:00 247.7 F 14.6
18:00 259.6 F 21.2
20:00 270.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.1 F 29.4
2:00 272.6 F 29.7
3:00 272.5 F 30.0
4:00 271.6 F 29.4

Min 271.5 F 29.3
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 233.5 F 0.0
10:00 215.3 F 0.0
15:00 216.6 F 0.0
20:00 221.4 F 0.0
25:00 223.4 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 21:34
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

12/16/2021 01:35 PM
ACCEPTED: X
REJECTED:
Operator: U02