

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001247  
11:07 06/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:19  
PRESSURIZING 3:24  
STERILIZING 4:42  
136.0 C 221kPa 4:42  
136.4 C 224kPa 5:12  
136.1 C 221kPa 5:42  
136.4 C 224kPa 6:12  
136.4 C 225kPa 6:42  
136.4 C 223kPa 7:12  
136.2 C 223kPa 7:42  
136.2 C 222kPa 8:12  
136.2 C 222kPa 8:12  
VENTING 8:12  
AIR DRYING 9:06  
CYCLE COMPLETE 24:06  
-----

Digital Signature #  
A593F6087EC0CE4A  
-----