
Midmark M11 - v1.0.0
Total Cycles: 2706

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2022 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.6 F 0.0
2:00 82.8 F 0.0
4:00 92.8 F 0.0
6:00 130.5 F 0.8
8:00 161.3 F 2.7
10:00 182.2 F 5.6
12:00 199.9 F 9.1
14:00 213.1 F 12.7
16:00 219.5 F 6.0
18:00 229.2 F 10.1
20:00 234.4 F 9.2
22:00 239.7 F 11.3
24:00 236.4 F 8.8
26:00 249.9 F 15.2
28:00 260.8 F 21.5
30:00 271.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.0
2:00 271.5 F 28.7
3:00 272.2 F 28.8
4:00 271.5 F 28.7

Min 271.3 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 225.4 F 0.0
10:00 225.3 F 0.0
15:00 229.6 F 0.0
20:00 230.0 F 0.0
25:00 226.5 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:39
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:38
CYCLE COMPLETE

12/21/2022 10:08 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____