
Midmark M11 - v1.0.5
Total Cycles: 2305

Name: RITTER2
Sterilizer ID: V2313228
04/27/2023 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.1 F 0.0
2:00 169.5 F 2.8
4:00 182.7 F 6.9
6:00 210.4 F 12.2
8:00 222.5 F 7.5
10:00 236.4 F 13.1
12:00 237.4 F 10.8
14:00 237.1 F 9.6
16:00 251.5 F 16.6
18:00 263.7 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.4 F 29.8
2:00 271.5 F 29.0
3:00 272.3 F 29.8
4:00 271.9 F 29.3

Min 271.1 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 222.1 F 0.0
10:00 228.4 F 0.0
15:00 226.8 F 0.0
20:00 226.1 F 0.0
25:00 225.5 F 0.0
30:00 227.0 F 0.0

Drying Complete

FILLING: 1:27
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

04/27/2023 11:19 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____