
Midmark M11 - v1.0.0
Total Cycles: 4027

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2023 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 72.3 F 0.1
4:00 84.2 F 0.9
6:00 116.2 F 2.9
8:00 148.6 F 5.5
10:00 172.1 F 8.4
12:00 186.3 F 11.5
14:00 207.3 F 3.3
16:00 212.9 F 5.2
18:00 221.7 F 8.3
20:00 231.4 F 12.0
22:00 232.6 F 7.5
24:00 242.8 F 12.4
26:00 239.9 F 10.1
28:00 251.9 F 15.9
30:00 262.0 F 21.9
32:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.2 F 28.8
2:00 271.0 F 28.2
3:00 272.3 F 28.9
4:00 271.3 F 28.2

Min 271.0 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 226.5 F 0.0
10:00 209.0 F 0.0
15:00 192.1 F 0.0
20:00 190.6 F 0.0
25:00 198.2 F 0.0
30:00 221.6 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 32:38
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:11:03
CYCLE COMPLETE

11/16/2023 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____