

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1477

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2022 01:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
2:00 179.8 F 3.5  
4:00 193.0 F 7.9  
6:00 219.5 F 13.5  
8:00 228.9 F 9.4  
10:00 236.8 F 9.1  
12:00 242.3 F 13.2  
14:00 242.7 F 12.4  
16:00 255.8 F 19.3  
18:00 267.5 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.6 F 30.1  
2:00 272.1 F 29.4  
3:00 271.2 F 29.2  
4:00 272.5 F 30.2  
  
Min 271.1 F 29.0  
Max 272.8 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.1 F 0.0  
5:00 224.7 F 0.0  
10:00 233.3 F 0.0  
15:00 225.3 F 0.0  
20:00 226.8 F 0.0  
25:00 226.8 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:17  
CYCLE COMPLETE

09/12/2022 02:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_