
Midmark M11 - v1.0.5
Total Cycles: 3957

Name: RITTER2
Sterilizer ID: V2313228
09/19/2024 09:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 73.9 F 0.0
4:00 79.4 F 0.0
6:00 98.4 F 0.1
8:00 147.8 F 0.3
10:00 175.4 F 0.7
12:00 194.8 F 3.3
14:00 205.7 F 6.1
16:00 216.5 F 9.5
18:00 227.3 F 13.1
20:00 229.3 F 8.3
22:00 239.7 F 12.8
24:00 238.5 F 10.4
26:00 236.9 F 8.9
28:00 249.4 F 15.2
30:00 260.2 F 21.3
32:00 270.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.5 F 29.3
2:00 271.5 F 28.8
3:00 272.5 F 29.3
4:00 271.6 F 28.6

Min 271.4 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 200.4 F 0.0
10:00 199.1 F 0.0
15:00 210.7 F 0.0
20:00 219.4 F 0.0
25:00 211.7 F 0.0
30:00 208.0 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:57
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:11:22
CYCLE COMPLETE

09/19/2024 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____