

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001469  
13:12 10/01/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:00  
PRESSURIZING 6:45  
STERILIZING 8:13  
136.0 C 220kPa 8:13  
135.6 C 218kPa 8:43  
136.3 C 217kPa 9:14  
136.4 C 223kPa 9:44  
136.4 C 221kPa 10:14  
136.1 C 220kPa 10:44  
135.8 C 218kPa 11:14  
136.4 C 219kPa 11:43  
VENTING 11:44  
AIR DRYING 12:52  
CYCLE COMPLETE 42:52

-----  
Digital Signature #  
40BB42E5524D637F  
-----