
Midmark M11 - v1.0.5
Total Cycles: 1904

Name: RITTER2
Sterilizer ID: V2313228
01/09/2023 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.3 F 0.0
2:00 78.2 F 0.0
4:00 104.6 F 0.0
6:00 122.3 F 0.0
8:00 146.1 F 0.1
10:00 174.8 F 0.2
12:00 196.1 F 3.3
14:00 213.4 F 6.6
16:00 222.8 F 10.2
18:00 229.9 F 9.1
20:00 236.2 F 10.4
22:00 235.4 F 8.6
24:00 245.0 F 12.9
26:00 244.2 F 12.6
28:00 256.0 F 18.9
30:00 266.7 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.6 F 29.4
2:00 271.5 F 28.6
3:00 272.7 F 29.5
4:00 271.9 F 28.8

Min 271.3 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 215.9 F 0.0
10:00 225.1 F 0.0
15:00 220.0 F 0.0
20:00 218.0 F 0.0
25:00 217.7 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:36
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

01/09/2023 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____