

-----  
Midmark M11 - v1.0.5  
Total Cycles: 157

Name: RITTER2  
Sterilizer ID: V2313228  
07/05/2021 08:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.5 F 0.0  
2:00 149.8 F 1.6  
4:00 152.5 F 4.5  
6:00 174.7 F 8.6  
8:00 198.1 F 13.4  
10:00 213.4 F 6.7  
12:00 227.6 F 11.6  
14:00 232.5 F 9.1  
16:00 238.9 F 10.4  
18:00 246.0 F 14.0  
20:00 258.0 F 20.4  
22:00 268.1 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.5  
1:00 271.9 F 30.0  
2:00 272.9 F 30.5  
3:00 272.9 F 30.3  
4:00 272.4 F 30.0

Min 271.2 F 29.3  
Max 273.3 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.1 F 0.0  
5:00 230.1 F 0.0  
10:00 230.2 F 0.0  
15:00 226.3 F 0.0  
20:00 216.6 F 0.0  
25:00 179.3 F 0.0  
30:00 169.9 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:28  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:40  
CYCLE COMPLETE

07/05/2021 09:13 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01