

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2015

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2022 09:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
2:00 207.3 F 3.6  
4:00 221.6 F 9.2  
6:00 230.4 F 6.8  
8:00 235.5 F 11.1  
10:00 234.7 F 8.1  
12:00 240.7 F 10.5  
14:00 245.8 F 13.2  
16:00 258.2 F 19.9  
18:00 269.4 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.4  
1:00 272.3 F 29.4  
2:00 272.7 F 29.3  
3:00 271.7 F 29.0  
4:00 272.9 F 29.7  
  
Min 271.0 F 28.3  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.4 F 0.0  
5:00 218.3 F 0.0  
10:00 230.0 F 0.0  
15:00 232.4 F 0.0  
20:00 227.1 F 0.0  
25:00 225.4 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 2:27  
HEATING: 19:07  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:47  
CYCLE COMPLETE

06/09/2022 10:48 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_