

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008410  
8:11 13/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:20  
PRESSURIZING 2:22  
STERILIZING 3:28  
136.4 C 223kPa 3:28  
136.4 C 223kPa 3:58  
136.4 C 221kPa 4:28  
136.2 C 222kPa 4:58  
136.4 C 220kPa 5:28  
136.4 C 222kPa 5:58  
136.2 C 222kPa 6:28  
136.4 C 219kPa 6:58  
136.4 C 219kPa 6:58  
VENTING 6:58  
AIR DRYING 7:49  
CYCLE COMPLETE 37:49

-----  
Digital Signature #  
544C84A4E0ED3E86  
-----