
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000232
13:02 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:40
STERILIZING 5:12
136.0 C 230kPa 5:12
136.4 C 231kPa 5:43
136.2 C 232kPa 6:13
136.3 C 232kPa 6:43
136.3 C 232kPa 7:13
136.3 C 232kPa 7:43
136.4 C 233kPa 8:13
136.4 C 232kPa 8:43
136.4 C 232kPa 8:43
VENTING 8:43
AIR DRYING 9:50
CYCLE COMPLETE 24:50

Digital Signature #
7DCE31006F3331C8
