
Midmark M11 - v1.0.0
Total Cycles: 939

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2021 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.1 F 0.0
2:00 187.9 F 2.4
4:00 192.0 F 6.2
6:00 210.5 F 10.7
8:00 221.9 F 5.7
10:00 231.0 F 9.9
12:00 235.9 F 9.2
14:00 242.3 F 12.3
16:00 240.8 F 10.8
18:00 253.0 F 16.9
20:00 263.6 F 23.2
22:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.6 F 29.2
2:00 271.9 F 28.6
3:00 271.5 F 28.7
4:00 272.5 F 29.2

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 211.2 F 0.0
10:00 219.9 F 0.0
15:00 224.2 F 0.0
20:00 246.3 F 0.0
25:00 235.6 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:40
CYCLE COMPLETE

06/24/2021 04:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____