
Midmark M11 - v1.0.0
Total Cycles: 3313

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2023 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 186.0 F 0.0
4:00 210.3 F 4.8
6:00 225.7 F 10.2
8:00 231.2 F 6.5
10:00 238.4 F 11.6
12:00 236.8 F 9.2
14:00 239.4 F 9.2
16:00 249.5 F 14.9
18:00 261.6 F 21.7
20:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.7 F 28.1
3:00 272.3 F 28.9
4:00 272.4 F 28.7

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 216.2 F 0.0
10:00 231.1 F 0.0
15:00 229.3 F 0.0
20:00 230.0 F 0.0
25:00 228.3 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

05/31/2023 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____