
Midmark M11 - v1.0.5
Total Cycles: 1156

Name: RITTER2
Sterilizer ID: V2313228
06/06/2022 07:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 77.4 F 0.0
4:00 81.4 F 0.0
6:00 103.1 F 0.1
8:00 138.5 F 3.0
10:00 165.8 F 6.4
12:00 186.9 F 10.2
14:00 208.8 F 7.8
16:00 221.3 F 6.9
18:00 231.2 F 11.0
20:00 232.2 F 7.4
22:00 241.8 F 12.4
24:00 240.0 F 10.7
26:00 252.0 F 16.7
28:00 262.5 F 22.8
30:00 272.1 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.7
2:00 271.6 F 28.8
3:00 271.9 F 29.5
4:00 272.3 F 29.4

Min 271.2 F 28.7
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 224.1 F 0.0
10:00 214.1 F 0.0
15:00 215.1 F 0.0
20:00 213.4 F 0.0
25:00 217.0 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:31
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:26
CYCLE COMPLETE

06/06/2022 08:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____