
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001133
14:17 22/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:34
STERILIZING 4:57
136.0 C 220kPa 4:57
136.3 C 224kPa 5:27
136.3 C 224kPa 5:57
136.4 C 225kPa 6:27
136.4 C 225kPa 6:57
136.2 C 222kPa 7:27
136.1 C 223kPa 7:57
136.3 C 223kPa 8:27
136.3 C 223kPa 8:27
VENTING 8:27
AIR DRYING 9:18
CYCLE COMPLETE 16:19

Digital Signature #
B0B6D74C9355DCB5
