

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2006

Name: RITTER 1  
Sterilizer ID: V2178198  
06/08/2022 07:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.3 F 0.0  
2:00 75.6 F 0.0  
4:00 96.4 F 1.3  
6:00 123.4 F 3.5  
8:00 149.7 F 6.5  
10:00 167.1 F 10.0  
12:00 209.5 F 10.9  
14:00 209.0 F 5.0  
16:00 219.5 F 8.4  
18:00 229.6 F 12.2  
20:00 230.9 F 7.4  
22:00 242.1 F 12.5  
24:00 238.9 F 10.0  
26:00 251.5 F 16.2  
28:00 262.1 F 22.3  
30:00 272.5 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.1 F 28.8  
2:00 272.2 F 29.2  
3:00 271.5 F 28.4  
4:00 272.4 F 29.1  
  
Min 271.2 F 28.4  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 203.4 F 0.0  
10:00 194.7 F 0.0  
15:00 195.3 F 0.0  
20:00 238.6 F 0.0  
25:00 238.6 F 0.0  
30:00 234.5 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:26  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:08:21  
CYCLE COMPLETE

06/08/2022 08:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_