
Midmark M11 - v1.0.5
Total Cycles: 1409

Name: RITTER2
Sterilizer ID: V2313228
08/24/2022 02:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.1 F 0.0
2:00 194.7 F 2.7
4:00 198.4 F 7.2
6:00 227.0 F 12.2
8:00 229.9 F 8.2
10:00 241.1 F 12.2
12:00 240.9 F 12.1
14:00 241.0 F 11.4
16:00 254.5 F 18.2
18:00 266.1 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.8
2:00 272.1 F 29.2
3:00 271.5 F 29.2
4:00 272.6 F 29.9

Min 271.4 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 231.7 F 0.0
10:00 244.4 F 0.0
15:00 236.5 F 0.0
20:00 226.4 F 0.0
25:00 225.0 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:40
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:38
CYCLE COMPLETE

08/24/2022 03:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____