

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001117  
14:12 09/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:48  
PRESSURIZING 2:58  
STERILIZING 4:01  
136.4 C 220kPa 4:01  
136.2 C 223kPa 4:31  
136.4 C 223kPa 5:01  
136.2 C 224kPa 5:31  
136.2 C 223kPa 6:01  
136.3 C 225kPa 6:31  
136.4 C 225kPa 7:01  
136.3 C 224kPa 7:31  
136.3 C 224kPa 7:31  
VENTING 7:31  
AIR DRYING 8:22  
CYCLE COMPLETE 15:18  
-----

Digital Signature #  
7DF295EF22B6F514  
-----