
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000584
16:26 12/08/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 4:27
STERILIZING 5:16
136.0 C 225kPa 5:16
136.2 C 229kPa 5:46
136.1 C 230kPa 6:16
135.7 C 229kPa 6:46
136.4 C 232kPa 7:16
135.8 C 230kPa 7:46
136.4 C 228kPa 8:16
136.3 C 232kPa 8:46
136.4 C 233kPa 9:16
136.4 C 231kPa 9:46
136.3 C 232kPa 10:16
136.4 C 233kPa 10:46
136.3 C 232kPa 11:16
136.3 C 232kPa 11:46
136.0 C 229kPa 12:16
136.4 C 232kPa 12:46
136.4 C 233kPa 13:16
136.3 C 228kPa 13:46
136.4 C 230kPa 14:16
136.3 C 232kPa 14:46
136.2 C 231kPa 15:16
136.2 C 231kPa 15:16
VENTING 15:16
AIR DRYING 16:03
CYCLE COMPLETE 31:03

Digital Signature #
CFF615A8CA2942B2
