
Midmark M11 - v1.0.5
Total Cycles: 1001

Name: RITTER2
Sterilizer ID: V2313228
04/19/2022 06:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 74.9 F 0.0
4:00 83.1 F 0.0
6:00 104.4 F 0.0
8:00 152.1 F 0.9
10:00 174.8 F 3.3
12:00 196.2 F 6.6
14:00 205.9 F 10.4
16:00 220.4 F 7.9
18:00 227.0 F 9.0
20:00 236.8 F 13.4
22:00 237.5 F 10.8
24:00 235.5 F 9.0
26:00 248.7 F 15.4
28:00 259.7 F 21.5
30:00 269.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.5 F 30.0
2:00 271.3 F 29.4
3:00 272.5 F 29.7
4:00 271.4 F 29.6

Min 271.2 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 210.2 F 0.0
10:00 203.3 F 0.0
15:00 200.4 F 0.0
20:00 206.9 F 0.0
25:00 214.5 F 0.0
30:00 207.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:57
CYCLE COMPLETE

04/19/2022 07:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____