
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001398
15:29 04/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:57
STERILIZING 4:37
136.5 C 224kPa 4:37
136.3 C 224kPa 5:07
136.3 C 224kPa 5:37
136.4 C 224kPa 6:07
136.4 C 224kPa 6:37
136.4 C 224kPa 7:07
136.4 C 223kPa 7:37
136.3 C 223kPa 8:07
136.3 C 223kPa 8:07
VENTING 8:07
AIR DRYING 9:04
CYCLE COMPLETE 9:06

Digital Signature #
96F3EF4608156F91
