
Midmark M11 - v1.0.5
Total Cycles: 2844

Name: RITTER2
Sterilizer ID: V2313228
09/07/2023 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.9 F 0.0
2:00 179.4 F 3.0
4:00 192.7 F 7.4
6:00 208.0 F 12.2
8:00 224.1 F 7.1
10:00 235.5 F 12.2
12:00 236.4 F 10.1
14:00 237.4 F 9.1
16:00 248.9 F 15.3
18:00 260.2 F 21.6
20:00 270.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.6 F 29.5
2:00 271.2 F 29.3
3:00 272.1 F 29.5
4:00 271.0 F 29.1

Min 271.0 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 219.0 F 0.0
10:00 226.5 F 0.0
15:00 213.5 F 0.0
20:00 206.8 F 0.0
25:00 216.9 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:46
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

09/07/2023 01:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____