
Midmark M11 - v1.0.0
Total Cycles: 2059

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2022 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.0 F 0.0
2:00 179.4 F 3.0
4:00 181.4 F 6.8
6:00 212.5 F 11.5
8:00 221.3 F 5.8
10:00 234.3 F 11.0
12:00 232.9 F 7.5
14:00 244.3 F 12.6
16:00 243.6 F 12.1
18:00 256.0 F 18.7
20:00 267.1 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.2
2:00 271.8 F 28.5
3:00 271.5 F 28.9
4:00 272.3 F 29.2

Min 271.2 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 219.4 F 0.0
10:00 215.0 F 0.0
15:00 226.1 F 0.0
20:00 239.3 F 0.0
25:00 237.0 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

06/27/2022 10:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____