
Midmark M11 - v1.0.5
Total Cycles: 1215

Name: RITTER2
Sterilizer ID: V2313228
06/22/2022 02:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 183.6 F 3.5
4:00 200.0 F 7.8
6:00 213.9 F 13.1
8:00 224.5 F 8.1
10:00 233.7 F 12.5
12:00 237.2 F 11.2
14:00 237.9 F 9.9
16:00 250.9 F 16.7
18:00 262.9 F 23.5
20:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.4 F 29.8
2:00 271.3 F 29.2
3:00 272.5 F 30.0
4:00 271.8 F 29.5

Min 271.3 F 29.2
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 221.1 F 0.0
10:00 222.5 F 0.0
15:00 229.1 F 0.0
20:00 224.1 F 0.0
25:00 221.3 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

06/22/2022 03:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____