
Midmark M11 - v1.0.0
Total Cycles: 3714

Name: RITTER 1
Sterilizer ID: V2178198
09/05/2023 08:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 72.5 F 0.0
4:00 101.0 F 0.0
6:00 132.4 F 0.0
8:00 141.1 F 1.4
10:00 167.8 F 3.2
12:00 190.9 F 5.4
14:00 201.2 F 8.7
16:00 212.8 F 12.4
18:00 222.3 F 5.9
20:00 232.3 F 10.0
22:00 236.6 F 9.7
24:00 241.4 F 11.5
26:00 237.7 F 9.4
28:00 250.8 F 15.8
30:00 261.8 F 22.0
32:00 271.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.1 F 28.6
2:00 272.8 F 29.1
3:00 272.0 F 28.8
4:00 271.9 F 28.5

Min 271.5 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 192.1 F 0.0
10:00 209.2 F 0.0
15:00 212.1 F 0.0
20:00 225.1 F 0.0
25:00 228.8 F 0.0
30:00 235.1 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:10:28
CYCLE COMPLETE

09/05/2023 09:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____