
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008182
10:06 19/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:38
STERILIZING 3:39
136.4 C 222kPa 3:39
136.2 C 223kPa 4:09
136.4 C 220kPa 4:39
136.4 C 223kPa 5:09
136.1 C 222kPa 5:39
136.2 C 223kPa 6:09
136.4 C 223kPa 6:39
136.4 C 222kPa 7:09
136.4 C 222kPa 7:09
VENTING 7:09
AIR DRYING 7:58
CYCLE COMPLETE 14:53

Digital Signature #
1D7CD30DA3E1B926
