

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 004246  
16:58 25/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:50  
PRESSURIZING 1:55  
STERILIZING 3:00  
136.4 C 223kPa 3:00  
136.4 C 225kPa 3:30  
136.2 C 223kPa 4:00  
136.2 C 222kPa 4:30  
136.4 C 225kPa 5:00  
136.4 C 220kPa 5:30  
136.4 C 225kPa 6:00  
136.2 C 222kPa 6:30  
136.2 C 222kPa 6:30  
VENTING 6:30  
AIR DRYING 7:17  
CYCLE COMPLETE 7:23

-----  
Digital Signature #  
9F363D9198EE9008  
-----