
Midmark M11 - v1.0.5
Total Cycles: 1261

Name: RITTER2
Sterilizer ID: V2313228
07/06/2022 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.0 F 0.0
2:00 129.2 F 1.5
4:00 149.8 F 4.3
6:00 172.1 F 8.2
8:00 196.0 F 12.8
10:00 213.3 F 6.6
12:00 226.6 F 11.4
14:00 230.0 F 7.7
16:00 240.9 F 13.1
18:00 240.9 F 11.5
20:00 252.9 F 17.7
22:00 263.6 F 24.2
24:00 272.0 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.5
1:00 272.3 F 30.2
2:00 272.1 F 29.7
3:00 271.2 F 29.5
4:00 272.1 F 30.3

Min 271.1 F 29.3
Max 272.4 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 198.6 F 0.0
10:00 217.7 F 0.0
15:00 224.2 F 0.0
20:00 231.0 F 0.0
25:00 232.0 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 24:13
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:24
CYCLE COMPLETE

07/06/2022 11:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____