
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000444
7:57 15/09/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:38
PRESSURIZING 4:38
STERILIZING 5:55
136.0 C 220kPa 5:55
136.3 C 224kPa 6:25
136.2 C 223kPa 6:55
136.2 C 223kPa 7:25
136.4 C 223kPa 7:55
136.2 C 221kPa 8:25
136.3 C 222kPa 8:55
136.3 C 223kPa 9:25
136.3 C 223kPa 9:25
VENTING 9:25
AIR DRYING 10:27
CYCLE COMPLETE 12:41

Digital Signature #
27111989EE600590
