
Midmark M11 - v1.0.5
Total Cycles: 247

Name: RITTER2
Sterilizer ID: V2313228
07/26/2021 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 178.3 F 3.0
4:00 184.7 F 7.1
6:00 204.4 F 11.9
8:00 218.8 F 6.5
10:00 231.3 F 11.3
12:00 232.4 F 8.5
14:00 241.0 F 11.7
16:00 244.9 F 13.3
18:00 257.0 F 19.6
20:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.7 F 29.6
2:00 271.5 F 28.9
3:00 272.7 F 29.8
4:00 272.1 F 29.2

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 192.3 F 0.0
10:00 207.1 F 0.0
15:00 196.1 F 0.0
Dry Cycle Aborted

FILLING: 1:45
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:18
DRYING: 17:12
TOTAL CYCLE: 00:47:40

07/26/2021 01:45 PM
ACCEPTED: X
REJECTED:
Operator: U01