

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002666  
8:33 08/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:57  
PRESSURIZING 2:01  
STERILIZING 3:06  
136.0 C 220kPa 3:06  
136.2 C 224kPa 3:36  
136.4 C 225kPa 4:06  
136.4 C 223kPa 4:36  
136.2 C 224kPa 5:06  
136.4 C 225kPa 5:36  
136.4 C 224kPa 6:06  
136.4 C 224kPa 6:36  
136.4 C 224kPa 6:36  
VENTING 6:36  
AIR DRYING 7:29  
CYCLE COMPLETE 7:53  
-----

Digital Signature #  
D2B37A315E7B1A50  
-----