

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002292  
10:38 19/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:08  
STERILIZING 3:12  
136.4 C 225kPa 3:12  
136.3 C 223kPa 3:42  
136.4 C 224kPa 4:12  
136.4 C 226kPa 4:42  
136.4 C 225kPa 5:12  
136.0 C 221kPa 5:42  
136.4 C 225kPa 6:12  
136.3 C 225kPa 6:42  
136.3 C 225kPa 6:42  
VENTING 6:42  
AIR DRYING 7:30  
CYCLE COMPLETE 37:30  
-----

Digital Signature #  
E3A63F5661B7231A  
-----