
Midmark M11 - v1.0.0
Total Cycles: 1633

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2022 09:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 167.8 F 1.3
4:00 180.8 F 4.4
6:00 196.7 F 8.8
8:00 218.4 F 8.7
10:00 229.3 F 8.1
12:00 241.0 F 13.3
14:00 240.1 F 10.6
16:00 236.3 F 8.5
18:00 251.2 F 15.5
20:00 263.0 F 22.4
22:00 271.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.8 F 29.1
2:00 271.6 F 28.3
3:00 272.8 F 28.8
4:00 272.0 F 28.7

Min 270.9 F 27.8
Max 273.1 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 222.1 F 0.0
10:00 236.4 F 0.0
15:00 227.3 F 0.0
20:00 220.5 F 0.0
25:00 220.0 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

02/15/2022 10:26 AM
ACCEPTED: X
REJECTED:
Operator: U02