

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001628  
10:02 29/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:29  
PRESSURIZING 2:34  
STERILIZING 3:39  
136.0 C 220kPa 3:39  
135.9 C 222kPa 4:09  
136.3 C 221kPa 4:39  
136.4 C 224kPa 5:09  
136.3 C 223kPa 5:39  
136.4 C 225kPa 6:09  
136.2 C 223kPa 6:39  
136.2 C 222kPa 7:09  
136.2 C 222kPa 7:09  
VENTING 7:09  
AIR DRYING 8:04  
CYCLE COMPLETE 8:09

-----  
Digital Signature #  
55453E26F7E819BF  
-----