
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002634
8:09 31/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:34
STERILIZING 3:44
136.2 C 222kPa 3:44
136.4 C 224kPa 4:14
136.3 C 220kPa 4:44
136.4 C 224kPa 5:14
136.3 C 220kPa 5:44
136.1 C 222kPa 6:14
136.4 C 222kPa 6:44
136.2 C 222kPa 7:14
136.2 C 222kPa 7:14
VENTING 7:14
AIR DRYING 8:05
CYCLE COMPLETE 8:42

Digital Signature #
641C95A5B37D823C
