

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1563

Name: RITTER 1  
Sterilizer ID: V2178198  
01/26/2022 07:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.0 F 0.0  
2:00 76.0 F 0.0  
4:00 87.5 F 0.3  
6:00 115.0 F 2.3  
8:00 143.8 F 5.1  
10:00 167.9 F 8.3  
12:00 186.4 F 11.8  
14:00 201.6 F 2.2  
16:00 213.8 F 5.5  
18:00 223.7 F 8.9  
20:00 232.5 F 12.7  
22:00 233.1 F 8.4  
24:00 244.0 F 13.3  
26:00 241.7 F 11.0  
28:00 253.2 F 16.8  
30:00 263.5 F 22.9  
32:00 272.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.6 F 29.0  
2:00 272.2 F 28.4  
3:00 271.5 F 28.3  
4:00 272.7 F 29.0  
  
Min 271.4 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 229.7 F 0.0  
10:00 225.3 F 0.0  
15:00 220.0 F 0.0  
20:00 217.2 F 0.0  
25:00 218.3 F 0.0  
30:00 220.9 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:20  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:10:25  
CYCLE COMPLETE

01/26/2022 09:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_