

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3566

Name: RITTER2  
Sterilizer ID: V2313228  
04/26/2024 03:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.1 F 0.0  
2:00 152.1 F 2.7

C103: HEATUP MODE  
STOP PRESSED  
CHAMBER TEMP 155F  
CHAMBER PRESS 3.7PSI  
\*\* ITEMS NOT STERILE \*\*

04/26/2024 03:21 PM