

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1674

Name: RITTER 1  
Sterilizer ID: V2178198  
02/25/2022 12:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 75.9 F 0.0  
4:00 87.4 F 0.2  
6:00 121.0 F 1.0  
8:00 150.0 F 2.0  
10:00 172.9 F 3.6  
12:00 189.9 F 6.2  
14:00 205.4 F 9.4  
16:00 217.5 F 12.9  
18:00 220.2 F 5.8  
20:00 230.9 F 9.8  
22:00 239.0 F 10.3  
24:00 240.5 F 11.3  
26:00 237.9 F 9.3  
28:00 250.7 F 15.6  
30:00 261.4 F 21.6  
32:00 271.3 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.5 F 28.9  
2:00 271.6 F 28.1  
3:00 272.2 F 29.0  
4:00 272.2 F 28.5  
  
Min 271.4 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
5:00 200.5 F 0.0  
10:00 208.5 F 0.0  
15:00 209.7 F 0.0  
20:00 208.9 F 0.0  
25:00 214.1 F 0.0  
30:00 221.2 F 0.0  
  
Drying Complete

FILLING: 1:02  
HEATING: 32:41  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:11:03  
CYCLE COMPLETE

02/25/2022 01:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_