
Midmark M11 - v1.0.0
Total Cycles: 4033

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2023 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 72.3 F 0.0
4:00 81.9 F 0.7
6:00 110.4 F 2.7
8:00 139.7 F 5.1
10:00 163.6 F 8.0
12:00 184.1 F 11.4
14:00 205.6 F 2.9
16:00 210.3 F 5.5
18:00 220.0 F 8.7
20:00 228.8 F 12.5
22:00 230.7 F 8.0
24:00 242.1 F 12.9
26:00 240.2 F 10.9
28:00 252.0 F 16.6
30:00 262.4 F 22.6
32:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.1
2:00 271.4 F 28.6
3:00 272.4 F 29.2
4:00 271.4 F 28.5

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 181.8 F 0.0
10:00 199.0 F 0.0
15:00 215.7 F 0.0
20:00 225.0 F 0.0
25:00 230.6 F 0.0
30:00 229.3 F 0.0

Drying Complete

FILLING: 1:01
HEATING: 32:30
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:11:08
CYCLE COMPLETE

11/17/2023 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____