
Midmark M11 - v1.0.0
Total Cycles: 3890

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2023 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 179.1 F 4.2
4:00 195.5 F 9.0
6:00 211.2 F 7.9
8:00 228.6 F 8.6
10:00 238.0 F 10.8
12:00 240.3 F 11.5
14:00 239.0 F 9.9
16:00 252.8 F 16.7
18:00 264.5 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.2 F 28.6
2:00 272.3 F 29.1
3:00 271.3 F 28.3
4:00 272.6 F 29.1

Min 271.1 F 28.3
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 248.7 F 0.0
10:00 245.4 F 0.0
15:00 234.5 F 0.0
20:00 227.2 F 0.0
25:00 227.7 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

10/11/2023 01:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____