
Midmark M11 - v1.0.0
Total Cycles: 883

Name: RITTER 1
Sterilizer ID: V2178198
06/10/2021 01:11 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.7 F 0.0
2:00 130.1 F 1.0
4:00 147.2 F 3.3
6:00 175.5 F 7.4
8:00 204.0 F 12.4
10:00 220.0 F 6.9
12:00 236.6 F 12.1
14:00 233.1 F 9.0
16:00 236.8 F 9.3
18:00 245.9 F 13.3
20:00 257.8 F 20.0
22:00 269.1 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.1 F 29.0
2:00 273.0 F 29.4
3:00 272.7 F 29.7

Min 271.3 F 28.9
Max 273.2 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.1 F 0.0
5:00 241.0 F 0.0
10:00 229.5 F 0.0
15:00 229.8 F 0.0
20:00 221.2 F 0.0
25:00 183.9 F 0.0
Dry Cycle Aborted

FILLING: 1:17
HEATING: 23:41
STERILIZING: 3:00
VENTING: 3:02
DRYING: 25:04
TOTAL CYCLE: 00:56:03

06/10/2021 02:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____