

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3730

Name: RITTER2  
Sterilizer ID: V2313228  
06/24/2024 09:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.7 F 0.0  
2:00 82.0 F 0.0  
4:00 80.8 F 0.0  
6:00 113.9 F 1.5  
8:00 140.6 F 4.0  
10:00 160.6 F 7.0  
12:00 189.9 F 10.7  
14:00 207.5 F 5.6  
16:00 217.3 F 6.8  
18:00 224.1 F 10.5  
20:00 231.5 F 8.5  
22:00 237.5 F 10.7  
24:00 234.4 F 8.2  
26:00 246.1 F 14.0  
28:00 257.1 F 19.9  
30:00 267.4 F 26.3  
32:00 272.4 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 272.0 F 29.4  
2:00 272.2 F 29.7  
3:00 272.3 F 29.9  
4:00 271.9 F 29.9  
  
Min 271.3 F 29.3  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.4 F 0.0  
5:00 215.4 F 0.0  
10:00 212.3 F 0.0  
15:00 198.1 F 0.0  
20:00 197.9 F 0.0  
25:00 200.3 F 0.0  
30:00 199.8 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:12  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:10:29  
CYCLE COMPLETE

06/24/2024 10:35 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_