
Midmark M11 - v1.0.0
Total Cycles: 2793

Name: RITTER 1
Sterilizer ID: V2178198
01/18/2023 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.7 F 0.0
2:00 166.1 F 2.6
4:00 180.1 F 6.2
6:00 201.0 F 10.8
8:00 214.1 F 3.9
10:00 227.0 F 8.8
12:00 237.3 F 12.0
14:00 238.9 F 10.7
16:00 235.5 F 8.6
18:00 250.3 F 15.4
20:00 261.8 F 22.0
22:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.2
2:00 271.7 F 28.4
3:00 271.9 F 28.9
4:00 272.5 F 29.1

Min 271.4 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 228.8 F 0.0
10:00 233.4 F 0.0
15:00 228.9 F 0.0
20:00 227.6 F 0.0
25:00 228.2 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

01/18/2023 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____