

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2461

Name: RITTER 1  
Sterilizer ID: V2178198  
10/20/2022 07:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.8 F 0.0  
2:00 79.3 F 0.0  
4:00 87.2 F 0.0  
6:00 114.7 F 1.9  
8:00 149.1 F 4.8  
10:00 171.5 F 8.1  
12:00 219.3 F 11.8  
14:00 215.5 F 3.7  
16:00 223.1 F 7.3  
18:00 233.5 F 11.1  
20:00 232.0 F 7.0  
22:00 240.5 F 11.6  
24:00 236.5 F 9.0  
26:00 249.4 F 15.0  
28:00 260.3 F 21.0  
30:00 270.0 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.8 F 29.1  
2:00 272.0 F 29.2  
3:00 272.4 F 28.9  
4:00 272.4 F 29.4

Min 271.4 F 28.7  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 197.3 F 0.0  
10:00 195.8 F 0.0  
15:00 193.4 F 0.0  
20:00 189.7 F 0.0  
25:00 187.7 F 0.0  
30:00 206.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 31:37  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:09:46  
CYCLE COMPLETE

10/20/2022 08:51 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02