
Midmark M11 - v1.0.5
Total Cycles: 3167

Name: RITTER2
Sterilizer ID: V2313228
11/29/2023 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.3 F 0.0
2:00 135.8 F 0.9
4:00 159.4 F 3.6
6:00 185.5 F 7.3
8:00 199.1 F 11.6
10:00 218.1 F 6.2
12:00 229.6 F 11.0
14:00 228.8 F 7.5
16:00 242.0 F 13.2
18:00 241.8 F 11.8
20:00 253.3 F 17.8
22:00 263.9 F 24.0
24:00 272.8 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.8
1:00 271.5 F 29.8
2:00 271.7 F 29.7
3:00 271.9 F 29.7
4:00 272.1 F 29.8

Min 271.5 F 29.7
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 210.4 F 0.0
10:00 195.1 F 0.0
15:00 197.6 F 0.0
20:00 196.0 F 0.0
25:00 194.4 F 0.0
30:00 192.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 25:03
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:03:46
CYCLE COMPLETE

11/29/2023 01:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____