

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5383

Name: RITTER 1  
Sterilizer ID: V2178198  
03/04/2025 09:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.3 F 0.0  
2:00 74.5 F 0.3  
4:00 93.6 F 1.6  
6:00 126.2 F 3.9  
8:00 144.8 F 6.5  
10:00 165.4 F 9.9  
12:00 186.0 F 13.5  
14:00 203.4 F 3.4  
16:00 214.8 F 6.5  
18:00 225.3 F 9.9  
20:00 230.8 F 10.5  
22:00 235.6 F 9.6  
24:00 239.0 F 9.9  
26:00 244.6 F 12.7  
28:00 256.1 F 18.7  
30:00 266.2 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.7 F 29.0  
2:00 272.1 F 28.8  
3:00 272.4 F 28.7  
4:00 272.3 F 28.9

Min 271.6 F 28.3  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 193.8 F 0.0  
10:00 194.5 F 0.0  
15:00 201.2 F 0.0  
20:00 204.2 F 0.0  
25:00 203.4 F 0.0  
30:00 204.5 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 31:42  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:10:00  
CYCLE COMPLETE

03/04/2025 10:39 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_