

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5251

Name: RITTER 1  
Sterilizer ID: V2178198  
01/16/2025 04:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.4 F 0.0  
2:00 112.4 F 0.0  
4:00 123.0 F 0.0  
6:00 160.8 F 0.4  
8:00 180.9 F 3.4  
10:00 206.7 F 7.0  
12:00 217.3 F 10.8  
14:00 224.6 F 6.6  
16:00 232.1 F 9.9  
18:00 237.4 F 9.5  
20:00 242.2 F 12.0  
22:00 239.5 F 10.3  
24:00 252.0 F 16.4  
26:00 262.7 F 22.5  
28:00 272.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 271.7 F 28.8  
2:00 272.4 F 28.8  
3:00 271.5 F 28.2  
4:00 271.9 F 28.9

Min 271.3 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 200.8 F 0.0  
10:00 193.7 F 0.0  
15:00 204.2 F 0.0  
20:00 227.1 F 0.0  
25:00 236.6 F 0.0  
30:00 229.0 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 28:20  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:06:59  
CYCLE COMPLETE

01/16/2025 05:40 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_